

INSPECTION CERTIFICATE




■ Meet DIN2270

- Jeweled bearing
- Carbide contact point
- Two measuring directions
 Supplied with one clamp: diameter Ø8mm
- Optional accessory: styli



PRECISION DIAL TEST INDICATOR (PROFESSIONAL TYPE)





Code	Range	Graduation	Accuracy	Hysteresis	Dial reading
2991-014*	0.14mm	0.001mm	3µm	2µm	0-70-0

^{*}Supplied with manufacturer inspection certificate