


RADIOGRAPHIC STEP TEST BLOCK

Unit: mm



note: groove depths from top to bottom are: 0.1, 0.2, 0.3, 0.4, 0.5, 0.7, 0.9

- Performance calibration of radiographic inspection equipment, determination of inspection process parameters, and verification of defect detection capability related to workpiece thickness

SPECIFICATION

Code	Material	Casing dimension (L×W×T)
RXT-C010	S355J0	150×80×20mm

STANDRD DELIVERY

Radiographic step block	1pc
Spacer block (10mm)	1pc
Spacer block (20mm)	1pc